


	<h1>Sampler Specification</h1>	Specification Number	675 8051H450S
		Revision	04

Part No. 675 8051H450S

Description Disposable Jumbo Pipette – 450mm

Material of Construction:

Tip (handle)	HDPE (virgin, FDA acceptable CFR 177.1520 & EU 10/2011)
Outer Tube	HDPE (virgin, FDA acceptable CFR 177.1520 & EU 10/2011)
Tip (product end)	HDPE (virgin, FDA acceptable CFR 177.1520 & EU 10/2011)



Method of Construction:	Ultrasonic welding
Nominal Length:	420mm
Diameter of Outer Tube:	21mm
Diameter of hole in tip:	Handle End (Blue) = 1.5mm Product End (Natural) = 3mm
Nominal Weight of Sampler:	50g
Assembly & Packing Environment:	Class 100,000 Medical Cleanroom
Individually Bagged?	Yes (heat sealed PE bag)
Method of Sterilisation:	Gamma Irradiation (25-45kGy)
Number of Samplers per Box:	50
BSE/TSE Free:	Yes
Recommended Storage Conditions:	Dry and ambient temperature