

	<h1 style="margin: 0;">Sampler Specification</h1>	Specification Number	675 8920H450S
		Revision	01

Part No. 675 8920H450S

Description SteriWare LiquiDip Mono – 450mm

Material of Construction: High Density Polyethylene (HDPE) - virgin

- Conforms to FDA CFR 177.1520
- Conforms to EU Regulations 10/2011
- Conforms to EC Regulations 1935/2004



Method of Construction: Ultrasonic welding

Overall Length 430mm

Maximum Diameter: 26mm

Outer Diameter of Tube: 21mm

Diameter of hole in tip: 6mm

Nominal Weight of Sampler: 50g

Assembly & Packing Environment: Class 100,000 Cleanroom

Individually Bagged? Yes (heat sealed PE bag)

Method of Sterilisation: Gamma Irradiation (25-45 kGy)

Number of Samplers per Box: 20

Recommended Storage Conditions: Dry and ambient temperature