

## QAQC UD DOSER Accurate Unit Dose Sampling

Ideal for accurately sampling specific volumes of liquids. Designed specifically for unit dose sampling of liquids in the pharmaceutical industry.

- Suitable for sampling low to medium viscosity liquids.
- Ideal for low volume (unit dose) sampling.
- Made from Sampler body 316 stainless steel, Syringe - PP & PE

Sample Chamber is designed so that the outside of the syringe is not contaminated during sampling. The syringe can therefore be used as a carrying container for taking the sample back to the laboratory.

## **Operation**



1. Select a new syringe



2. Connect the syringe to the push rod



3. Attach outer cover. The sampler is now ready for use.

Tel: (804) 318-3686



4. Pull up handle to take the sample, the graduations allow an accurate sample to be taken

| 5 ml Doser          |           | 30 ml Doser          |        | 60 ml Doser          |        |
|---------------------|-----------|----------------------|--------|----------------------|--------|
| Stock No.           | Length    | Stock No.            | Length | Stock No.            | Length |
| 675 5010A600        | 600mm     | 675 5010B600         | 600mm  | 675 5010C600         | 600mm  |
| 675 5010A120        | 00 1200mm | 675 5010B1200        | 1200mm | 675 5010C1200        | 1200mm |
| 675 5010A180        | 00 1800mm | 675 5010B1800        | 1800mm | 675 5010C1800        | 1800mm |
|                     |           |                      |        |                      |        |
| Spare 5 ml Syringes |           | Spare 30 ml Syringes |        | Spare 60 ml Syringes |        |

| 675 5010S05 Pack of 100 675 5010S30 Pack of 25 675 5010S60 Pack of 2 | Spare 5 m   | l Syringes  | Spare 30 ml | Syringes   | Spare 60 ml S | yringes    |
|--|-------------|-------------|-------------|------------|---------------|------------|
|  | 675 5010S05 | Pack of 100 | 675 5010S30 | Pack of 25 | 675 5010S60   | Pack of 25 |

| Accessories | Cleaning brushes See page 130 |              |  |  |  |
|-------------|-------------------------------|--------------|--|--|--|
|             | Storage cases                 | See page 131 |  |  |  |
|             | XRF Certification             | See page 140 |  |  |  |

